

I. INTENDED USE:

The UTAK Validity Control 3 is for use as quality control materials that will help identify substituted, diluted, and adulterated urine samples. The control material will generate data that checks and evaluates a test method. In the case of screening, it is important to have a control material that contains the desired analyte at or near the desired cut off value so that a continuous quality control program is obtained.

II. SUMMARY AND PRINCIPLES:

Several different techniques are used for evaluating or estimating the variance of results. The three subjects summarized below must be considered with any test method.

1. PREVENTIVE MEASURES:

These measures are usually contained in the design of the test method and include consideration for reagents, equipment, and operator errors. These measures are designed to minimize variance.

2. QUALITY CONTROL MEASURES:

When a quality control sample is analyzed at the same time and in the same manner as a patient specimen, an estimate of variance is obtained for the test method. This estimate of variance can be compared to the acceptable limits of variance of the test method.

3. STATISTICAL ANALYSIS OF PATIENT RESULTS:

As an aid in evaluating overall test results, the past experience of expected results can be compared to the results of any given test run. For example, it would not be expected that all results of a given test run be in an elevated range.

Quality control materials are widely used as a means to aid in the evaluation of test results. The following subjects are to be considered in the use of any control material.

- | | |
|-----------------|----------------------------|
| 1. Multi-Level | NORMAL / ELEVATED |
| 2. Matrix | HUMAN / ANIMAL / CHEMICAL |
| 3. Availability | SUFFICIENT FOR STATISTICS |
| 4. Form | LIQUID / FROZEN / DRIED |
| 5. Variety | DIFFERENT THAN CALIBRATORS |

The UTAK Validity Control 3 will generate data that check and evaluate the results of a test method over the low, normal, and elevated ranges. The principles of statistics require that the same material be available for comparison for any given time period. Statistical accuracy requires that a test method be defined for variance and be calibrated with a suitable standard. The quality control materials that are used must be of a sufficient variety so that the measurements and the data that are obtained are independent of the calibration standards. By using a variety of materials, the entire test method can be continuously evaluated to insure reliable results.

III. PRODUCT DESCRIPTION:

The matrix for the UTAK Validity Control 3 Control is prepared from distilled water. The levels of the Adulterants, Creatinine, Specific Gravity, and pH are adjusted to the desired range for each lot prepared. Quality control before, during, and after the preparation of the control material insures that each lot is comparable and of the same high quality. The listed ranges were verified by analytical methods similar to those actually used in a testing laboratory.

IV. PRECAUTIONS:

- For in vitro diagnostic use only.
- For analytical use only.

V. STORAGE AND STABILITY:

- Store liquid control material at 2-8°C (35-46°F). Stable to expiration date printed on the label.

VI. PROCEDURE:

- Swirl gently 3-4 minutes to ensure a homogeneous mixture.
- Swirl gently each time an aliquot is removed to ensure a homogeneous mixture.
- Assay control material in same manner as patient specimens, following the exact same instructions from the entire test method.
- Record the results obtained on a quality control chart that describes statistical limits for the test method and the particular lot of control material.

VII. LIMITATIONS:

- Control material is for use in quality control programs only; it is not intended for use as a calibration standard.
- Check the lot number on each vial to be sure it corresponds to the lot number printed on the insert.
- Control material approximates a patient specimen.

VIII. EXPECTED VALUES:

- Laboratories should establish their own statistical values for precision and expected ranges.

Lot Number : C4565		Expiration Date : 2022-01-31
Analyte	Target Value	Units
Chromium VI	n/a	µg/mL
Creatinine	3.0 – 4.0	mg/dL
Nitrites	200 – 250	µg/mL
Oxidants	n/a	µg/mL
pH	10.0 – 10.8	n/a
Specific Gravity	0.9950 – 1.0005	n/a

UTAK's express and implied warranties (including merchantability and fitness) are conditioned on the observance of UTAK's insert directions with respect to the use of UTAK's products.

For technical assistance call: UTAK Technical Service (888) 882-5522

UTAK Laboratories, Inc.
25020 Avenue Tibbitts
Valencia, CA 91355
USA
T: (888) 882-5522
F: (661) 294-9272
E: wlovecontrol@utak.com

Product Number: 17012
Product Description: Validity Control 3
Lot Number: C4565
Expiration Date: 01/22
Volume: 1x25mL Vials, Aqueous

EC Authorized Representative
Emergo Europe
Prinsessegracht 20
2514 AP, The Hague
The Netherlands

