

FOUR – LEVEL LEAD

WHOLE BLOOD TOXICOLOGY CONTROL

I. INTENDED USE:

Many analytes can be monitored in patients by using analytical test methods. The UTAK Four – Level Lead Control is for use as a quality control material for monitoring the accuracy and precision of procedures that measure the levels of lead in blood. It is intended for use on a continuous basis so that the deviations of reagents and analytical instrumentation can be detected on a statistical basis.

II. SUMMARY AND PRINCIPLES:

Several different techniques are used for evaluating or estimating the variance of results. The three subjects summarized below must be considered with any test method.

1. PREVENTIVE MEASURES:

These measures are usually contained in the design of the test method and include consideration for reagents, equipment, and operator errors. These measures are designed to minimize variance.

2. QUALITY CONTROL MEASURES:

When a quality control sample is analyzed at the same time and in the same manner as a patient specimen, an estimate of variance is obtained for the test method. This estimate of variance can be compared to the acceptable limits of variance of the test method.

3. STATISTICAL ANALYSIS OF PATIENT RESULTS:

As an aid in evaluating overall test results, the past experience of expected results can be compared to the results of any given test run. For example, it would not be expected that all results of a given test run be in an elevated range.

Quality control materials are widely used as a means to aid in the evaluation of test results. The following subjects are to be considered in the use of any control material.

- | | |
|-----------------|----------------------------|
| 1. Multi-Level | THERAPEUTIC / TOXIC |
| 2. Matrix | HUMAN / ANIMAL / CHEMICAL |
| 3. Availability | SUFFICIENT FOR STATISTICS |
| 4. Form | LIQUID / FROZEN / DRIED |
| 5. Variety | DIFFERENT THAN CALIBRATORS |

The UTAK Four – Level Lead Control is prepared from normal human materials and will generate data that checks and evaluates the results of a test method over the normal and elevated ranges. The principles of statistics require that the same material be available for comparison for any given time period. Dried control materials both extend the usable time period and allow larger quantities to be available. Statistical accuracy requires that a test method be defined for variance and be calibrated with a suitable standard. The quality control materials that are used must be of a sufficient variety so that the measurements and the data that are obtained are independent of the calibration standards. By using a variety of materials, the entire test method can be continuously evaluated to insure reliable results.

III. PRODUCT DESCRIPTION:

The matrix for the UTAK Four – Level Lead Control is prepared from normal human whole blood. Lead is added and adjusted to the desired concentration range for each lot prepared. Quality control before, during, and after the preparation of the control material ensures that each lot is of the same quality.

IV. PRECAUTIONS:

- Although the whole blood has been tested and found negative for HBsAg by RIA and HIV by EIA, the control material should be treated as any other potentially infectious agent.
- For in vitro diagnostic use only.
- For analytical use only.

V. STORAGE AND STABILITY:

- Store dried control material at 2-8°C (35-46°F). Stable to the expiration date printed on the insert and label.
- Store reconstituted control material at 2-8°C (35-46°F). **Stable for 30 days after reconstitution.**

VI. PROCEDURE:

- Remove cap from each vial to be used.
- Reconstitute control material by adding exactly **3.0 mL of distilled water**, using a **3 mL** volumetric pipette or equivalent. **To avoid contamination of control material use metal free pipette and dilutors.**
- Replace cap and swirl gently by hand 5-10 minutes. **Do not use a rotary or rocker mixer.**
- Allow control material to equilibrate for one hour at a room temperature of 18-25°C (64-77°F). Continue to swirl gently by hand until a homogeneous mixture is attained.**
- Mix thoroughly each time an aliquot is removed to ensure a homogeneous mixture.
- Assay control material (recommended in duplicate) in same manner as patient specimens, following the exact same instructions from the entire test method.
- Record the results obtained on a quality control chart that describes the statistical limits for the test method and the particular lot of control material.

VII. LIMITATIONS:

- Control material is for use in quality control programs only; it is not intended for use as a calibration standard.
- Check the lot number on each vial; be sure it corresponds to the lot number printed on the insert.
- Results are dependent upon proper storage, reconstitution accuracy, and adequate mixing.
- Control material approximates a patient specimen; it has not been assayed for any analytes not listed in the tables below.

VIII. EXPECTED VALUES:

- Listed in the table below is the *Reference Value*. The *Reference Value* is derived from replicate analysis performed by an independent reference laboratory testing for lead.
- The *Reference Value* is determined by Graphite Furnace Atomic Absorption Spectroscopy (GFAAS), and Inductively Coupled Plasma / Mass Spectrometry (ICP/MS).
- Laboratories should establish their own mean values; the mean of several determinations may not duplicate the values listed below, but should fall within ±15% of the *Reference Value*.

Lead Level I Product # 93311		Lot Number : A6265 Expiration Date : 11/18		
Method	Metric Units		Standard International Units	
	Units	Reference Value	Units	Reference Value
ICP / MS GFAAS	µg/dL	8	µmol/L	0.386
	µg/dL	6.6	µmol/L	0.319

Lead Level II Product # 93312		Lot Number : A6266 Expiration Date : 11/18		
Method	Metric Units		Standard International Units	
	Units	Reference Value	Units	Reference Value
ICP / MS GFAAS	µg/dL	20.4	µmol/L	0.987
	µg/dL	19.7	µmol/L	0.951

Lead Level III Product # 93313		Lot Number : A6267 Expiration Date : 11/18		
Method	Metric Units		Standard International Units	
	Units	Reference Value	Units	Reference Value
ICP / MS GFAAS	µg/dL	38.2	µmol/L	1.84
	µg/dL	37.0	µmol/L	1.79

Lead Level IV Product # 93314		Lot Number : A6268 Expiration Date : 11/18		
Method	Metric Units		Standard International Units	
	Units	Reference Value	Units	Reference Value
ICP / MS GFAAS	µg/dL	59.4	µmol/L	2.87
	µg/dL	55.5	µmol/L	2.68

UTAK's express and implied warranties (including merchantability and fitness) are conditioned on the observance of UTAK's insert directions with respect to the use of UTAK's products.

For technical assistance call: **UTAK Technical Service (800) 235-3442**

UTAK LABORATORIES, INC.
25020 AVENUE TIBBITTS
VALENCIA, CA 91355
TEL: (661) 294-3935
FAX: (661) 294-9272
E-MAIL: INQUIRIES@UTAK.COM

PRODUCT NUMBER:
93311, LEVEL I
93312, LEVEL II
93313, LEVEL III
93314, LEVEL IV
DFID: LWBTE
5x3ML VIALS, DRIED

EC AUTHORIZED REPRESENTATIVE
EMERGO EUROPE
MOLENSTRAAT 15
2513 BH, THE HAGUE
THE NETHERLANDS

