

Certificate of Analysis

Product Information

| Product Name: | Blank Whole Blood |
|----------------------|---|
| Product Number: | 44600-WB(F) |
| Date of Manufacture: | 11/22 |
| Lot Number: | D2428 |
| Expiration Date: | 2024-11-30 |
| Intended Use: | In vitro diagnostic |
| Homogeneity: | Sufficiently homogeneous for analytical testing |

Precautions

| Matrix: | Whole Blood |
|-------------------|--|
| Form: | Frozen |
| Anticoagulant: | Potassium Oxalate |
| Preservative: | Sodium Chloride |
| Preservative: | Sodium Fluoride |
| pH: | 6-7 |
| Virology: | Non-reactive for HBsAg, HCV, and HIV |
| Microbial Growth: | <100 colony-forming units per milliliter |

Storage and Stability

| Storage Condition: | -10°C (14°F) | | |
|--|--------------|--|--|
| Open Vial Stability: | 30 days* | | |
| Vial Volume: | 250 mL | | |
| *The matrix may be thawed, aliquoted, and refrozen; although, this product should not be exposed to multiple | | | |

The matrix may be thawed, aliquoted, and refrozen; although, this product should not be exposed to multiple freeze/thaw cycles.

ELISA Raw Material Screen

UTAK Laboratories, Whole Blood Pool 22MCTA3206 - TC24143

| Analyte | Qualitative Value | Cutoff Value | Unit |
|---|-------------------|--------------|-------|
| d-Amphetamine | Negative | 5 | ng/mL |
| d-Methamphetamine | Negative | 5 | ng/mL |
| Benzoylecgonine (BE) | Negative | 5 | ng/mL |
| Dextromethorphan | Negative | 10 | ng/mL |
| Diphenhydramine | Negative | 10 | ng/mL |
| Ethanol | Negative | 20 | mg/dL |
| Morphine | Negative | 5 | ng/mL |
| Oxazepam | Negative | 10 | ng/mL |
| Phencyclidine (PCP) | Negative | 2 | ng/mL |
| Secobarbital | Negative | 5 | ng/mL |
| (l)-9-Carboxy-11-nor-Delta-9-THC (THC-COOH) | Negative | 1 | ng/mL |





UTAK Certificate of Analysis

Quality Assurance

